
# **AMMONIA**



#### 1. PERFORMANCE

1) Measuring range 0.5-10%Number of pump strokes  $1(100 \text{m} \ell)$ 

2) Sampling time : 1.5 minutes/1 pump stroke

3) Detectable limit 0.04%4) Shelf life 3 years 5) Operating temperature  $0 \sim 40\%$ 

6) Reading : Direct reading from the scale calibrated by 1 pump stroke

7) Colour change : Pink→Gray · Yellow

### 2. RELATIVE STANDARD DEVIATION

RSD-low: 5% RSD-mid.: 5% RSD-high: 5%

### 3. CHEMICAL REACTION

By reacting with Phosphoric acid, PH indicator is discoloured.  $NH_3 + H_3PO_4 \rightarrow (NH_4)_2HPO_4$ 

### 4. CALIBRATION OF THE TUBE

STANDARD GAS CYLINDER METHOD

## 5. INTERFERENCE AND CROSS SENSITIVITY

Substance	Interference	Coexistence
Amines	Similar stain is produced.	Higher readings are given.