NICKEL CARBONYL



1. PERFORMANCE

1) Measuring range 20-700 ppmNumber of pump strokes $1 (100 \text{m} \ell)$

2) Sampling time : 3 minutes/1 pump stroke

3) Detectable limit : 10 ppm 4) Shelf life : 6 months 5) Operating temperature : $0 \sim 40 \, ^{\circ}$ C

6) Reading : Concentration chart method 7) Colour change : Pale yellow→Dark purple

2. RELATIVE STANDARD DEVIATION

RSD-low: 10% RSD-mid.: 10% RSD-high: 5%

3. CHEMICAL REACTION

By reacting with Gold chloride (III), colloidal gold is liberated. Ni (CO)₄ + AuCI₃→ Au

4. CALIBRATION OF THE TUBE

STANDARD GAS GYLINDER METHOD

5. INTERFERENCE AND CROSS SENSITIVITY

Substance	Interference	ppm	Coexistence
Arsine	Similar stain is produced.	10	Higher readings are given.
Iron carbonyl	"	10	"
Mercury vapour	"	10	"
Hydrogen sulphide	Brown stain is produced.	10	"
Sulphur dioxide	Pale blue stain is produced.	10	"
Carbon monoxide	The accuracy of readings is not affected.	1,000	"
Acetylene	"	3%	"

