



1. PERFORMANCE


- | | | |
|--------------------------|---|-----------|
| 1) Measuring range | : 0.2-2.0 % | 0.1-0.2 % |
| Number of pump strokes | 1 (100mℓ) | 2 (200mℓ) |
| 2) Sampling time | : 3.5 minutes/1 pump stroke | |
| 3) Detectable limit | : 0.08 % (200mℓ) | |
| 4) Shelf life | : 2 years | |
| 5) Operating temperature | : 0 ~ 40℃ | |
| 6) Reading | : Direct reading from the scale calibrated by 1 pump stroke | |
| 7) Colour change | : White → Dark brown | |

2. RELATIVE STANDARD DEVIATION

RSD-low : 10% RSD-mid. : 10% RSD-high : 10%

3. CHEMICAL REACTION

Iodine pentoxide is reduced.



4. CALIBRATION OF THE TUBE

GAS CHROMATOGRAPHY

5. INTERFERENCE AND CROSS SENSITIVITY

Substance	Interference	%	Coexistence
Trichloroethylene	Yellow stain is produced.	0.2	Higher readings are given.
1,1,1-Trichloroethane	Orange stain is produced.	0.3	∕
1,2-Dichloroethylene	∕	0.1	∕
Vinyl chloride	∕	0.02	∕
Aromatic hydrocarbons	Blackish brown stain is produced.	—	The accuracy of readings is not affected.
Carbon monoxide	Brownish-red stain is produced.	0.05	Higher readings are given.
Carbon tetrachloride	The accuracy of readings is not affected.		The accuracy of readings is not affected.